

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0805 Accredited to ISO/IEC 17025:2017	Testo Industrial Services Ltd Issue No: 031 Issue date: 26 January 2026	
	Stanley House Old Brick Kiln Monk Sherborne Road Ramsdell Hampshire RG26 5PR	Contact: Mr Paul Miller Tel: +44 (0) 1256 648989 E-Mail: info@testotis.co.uk Website: www.testotis.co.uk
Calibration performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Stanley House Old Brick Kiln Monk Sherborne Road Ramsdell Hampshire RG26 5PR	Humidity Temperature Air velocity Pressure Torque	P

Site activities performed away from the location listed above:

Location details	Activity	Location code
Any customer premises The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	Temperature Humidity Pressure Profiling of humidity and temperature chambers	S



0805
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 031 Issue date: 26 January 2026

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
HUMIDITY Relative humidity	For the temperature range 0 °C to 20 °C	0.11 °C	Calibration by comparison with a reference chilled mirror hygrometer and thermometers Calibration of devices with an electrical output may be undertaken	P
	10 %rh to 35 %rh	0.60 %rh		
	35 %rh to 65 %rh	1.05 %rh		
	65 %rh to 95 %rh	1.35 %rh		
	For the temperature range 20 °C to 30 °C	0.11 °C		
	10 %rh to 40 %rh	0.55 %rh		
	40 %rh to 70 %rh	0.90 %rh		
	70 %rh to 80 %rh	0.97 %rh		
	80 %rh to 95 %rh	1.10 %rh		
	For the temperature range 30 °C to 70 °C	0.11 °C		
	10 %rh to 35 %rh	0.45 %rh		
	35 %rh to 65 %rh	0.80 %rh		
65 %rh to 95 %rh	1.00 %rh			
Relative humidity	For the temperature range 5 °C to 10 °C	0.49 °C	Calibration by comparison with a reference hygrometer and thermometers	P & S
	10 %rh to 20 %rh	1.32 %rh		
	20 %rh to 50 %rh	1.91 %rh		
	50 %rh to 80 %rh	2.85 %rh		
	For the temperature range 10 °C to 20 °C	0.38 °C		
	10 %rh to 20 %rh	1.24 %rh		
	20 %rh to 50 %rh	1.55 %rh		
	50 %rh to 80 %rh	2.20 %rh		
	80 %rh to 95 %rh	2.41 %rh		
	For the temperature range 20 °C to 30 °C	0.40 °C		
	10 %rh to 20 %rh	1.18 %rh		
	20 %rh to 50 %rh	1.45 %rh		
	50 %rh to 80 %rh	1.63 %rh		
	80 %rh to 95 %rh	2.21 %rh		
	For the temperature range 30 °C to 40 °C	0.69 °C		
	10 %rh to 20 %rh	1.29 %rh		
	20 %rh to 50 %rh	2.09 %rh		
	50 %rh to 80 %rh	2.86 %rh		



0805
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 031 Issue date: 26 January 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
HUMIDITY continued				
Dew point/Frost point	-30 °C to +69 °C	0.16 °C	Calibration by comparison with a reference chilled mirror hygrometer	P
Profiling of humidity and temperature chambers	+20 °C to +85 °C 20 %rh to 95 %rh	3.30 %rh	Calibration by comparison with a reference chilled mirror hygrometer and temperature probes Uncertainty achieved will depend on the performance of the chamber at the time of calibration.	S
Saturated salt capsules	Performed at 25 °C 6 %rh to 35 %rh 35 %rh to 76 %rh 76 %rh to 90 %rh	1.0 %rh 1.4 %rh 1.7 %rh		P
TEMPERATURE			Calibration by comparison with a reference instrument	
Electronic thermometers with indicators and data-loggers	-90 °C to 0 °C 0 °C to +110 °C +110 °C to +200 °C	0.020 °C 0.038 °C 0.041 °C	In a liquid bath	P
	-70 °C to +70 °C 70 °C to 180 °C	0.11 °C 0.13 °C	In an air chamber	
Resistance thermometers	-90 °C to 0 °C 0 °C to +110 °C +110 °C to +200 °C	0.043 °C 0.054 °C 0.056 °C	In a liquid bath	P
	-70 °C to +45 °C +45 °C to +180 °C	0.18 °C 0.20 °C	In an air chamber	
Electronic thermometers with indicators and data-loggers	-80 °C to +180 °C	0.26 °C	In customers environment	S
Temperature-controlled ovens, environmental chambers, fridges/refrigerators, freezers (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	-40 °C to +200 °C	0.90 °C	Single and multipoint time-dependent temperature profiling, also referred to as spatial temperature surveying or mapping	S



0805
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 031 Issue date: 26 January 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
AIR VELOCITY			Calibration by comparison with a reference in a characterised wind tunnel.	
Calibration of vane anemometers by comparison	0.2 m/s to 1 m/s 1 m/s to 15 m/s 15 m/s to 20 m/s 20 m/s to 30 m/s	0.10 m/s 0.20 m/s 0.32 m/s 0.60 m/s	Anemometers up to 100 mm in diameter can be calibrated.	P
Calibration of thermal anemometers by comparison	0.1 m/s to 1 m/s 1 m/s to 10 m/s 10 m/s to 15 m/s 15 m/s to 20 m/s	0.13 m/s 0.37 m/s 0.75 m/s 0.91 m/s	Uncertainty and range achieved will depend on the size of the anemometer calibrated	
PRESSURE			Methods consistent with EURAMET CG17.	
Gas pressure gauge				
Calibration of pressure indicating instruments and gauges	-85 kPa to -2 kPa -2 kPa to -200 Pa -200 Pa to 0 Pa 0 Pa to 200 Pa 200 Pa to 1 kPa 1 kPa to 100 kPa 100 kPa to 7 MPa	0.035 % + 5.3 Pa 17 Pa 4.9 Pa 1.4 Pa 0.015% + 6.3 Pa 0.035 % + 5.3 Pa 0.013 % + 1.0 kPa	Absolute pressure calibrations can be undertaken using associated barometric pressure measurement correction. The uncertainties quoted will be increased by 72 Pa	P & S
Gas pressure absolute			Calibrations of pressure devices with an electrical output may be undertaken	
Calibration of pressure indicating instruments and gauges	75 kPa to 115 kPa	71 Pa		
TORQUE				
Hand Torque Tools (Including Drivers)	To BS EN ISO 6789-2:2017 0.2 N·m to 1000 N·m See notes 1 and 2	1.0 %	1. The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated	P
	To BS EN ISO 6789-2:2003 0.2 N·m to 1000 N·m	1.0 %	2. Calibrations may be given in other torque units lbf.in, lbf.ft, oz.in	
END				



0805
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 031 **Issue date:** 26 January 2026

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$