

CERTIFICATE OF CALIBRATION

Issued by

TESTO LIMITED

DATE OF ISSUE 12 March 2020

CERTIFICATE NUMBER Test123



0805

Page 1 of 2 pages



Testo Limited
Newman Lane, Alton
Hampshire, GU34 2QJ
Tel: 01420 544433
Fax: 01420 544419

Approved Signatory

Name xxx

Signature

Customer name	Testo Limited Newman Lane Alton GU34 2QJ
Order number	None
Customer reference	Unmarked
Description	Testo 417
Serial number	123456789
Condition	Satisfactory
Calibrated range	0.5 to 20 m/s
Date of calibration	12 March 2020
Date received	12 March 2020

The instrument was calibrated, in the centre of a low turbulence wind tunnel and comparisons were made against a transfer standard.

At each applied velocity a period of 1 minute was allowed for stabilization before any measurements were taken. Average values were determined from readings recorded over 30 seconds; this was repeated three times for both the reference standard and the UUT. The resultant velocity readings shown are the average of the three measurements.

Calibration performed by:

xxx

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number

Test123

Page 2 of 2 pages

UKAS Accredited Calibration Laboratory number 0805

As found results

The UUT was positioned with the airflow in the direction as marked by the arrow on the instrument.

Laboratory conditions :

Atmospheric pressure: 0 hPa \pm 1 hpa

Temperature: 0 °C \pm 2 °C

Humidity: 0 %rh \pm 10 %rh

Applied velocity m/s	Indicated velocity m/s	Error m/s	Uncertainty \pm m/s
0.50	0.50	0.00	0.20
5.00	5.00	0.00	0.21
10.00	10.00	0.00	0.29
15.00	15.00	0.00	0.34
20.00	20.00	0.00	0.34

Results above are only applicable to the instrument tested.

An additional uncertainty of ± 0.01 m/s is also applicable for the resolution of the digital display. The uncertainties refer to the measurement and are not intended to indicate the specification or repeatability of the instrument.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.